

Product Datasheet **PerCP/Cyanine5.5 Anti-Human CD1a Antibody [OKT-6]** Catalogue Code: AGEL2251

Antibody Data

Product SKU:	AGEL2251	Clone:	OKT-6
Applications:	FCM		
Reactivity:	Human		

Important Note:

Centrifuge before opening to ensure complete recovery of vial contents.

Product Information:

Alternate Names: Uniprot ID:	T6; R4;CD 1a;CD1A ;T-cell surface glycoprotein CD1a; P06126			
Background:	CD1a is a 49 kD member of the immunoglobulin superfamily also known as T6 and R4. It is a type I membrane glycoprotein with structural similarities to MHC class I and is non- covalently associated with β 2-microglobulin. CD1a plays a role in non-peptide glycolipid antigen presentation to CD1-restricted T cells. It is expressed on cortical double positive and single positive thymocytes, Langerhans cells, and dendritic cells. In addition to antigen presentation, CD1a has been implicated in thymic T cell development.			
Form:	Liquid	PerCP/Cyanine5.5 Excitation and Emission Spectra		
Conjugation:	PerCP/Cyanine 5.5	100		
Size:	20 Tests, 100 Tests, 200 Tests			
Host Species:	Mouse	Normalized (%)		
Isotype:	Mouse IgG1, κ	20 0 350 400 450 500 550 600 650 700 750 800 850 Wavelength (nm)		
		Ex:440;480;675 nm; Em:675 nm		
		0 350 400 450 500 550 600 650 700 750 800 850		

Isotype Control: PerCP/Cyanine5.5 Mouse IgG1, κ Isotype Control[MOPC-21] [Product AGEL2251]

Storage Buffer: Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer and 1% protein protectant.

Shipping: Biological ice pack at 4°C



- **Stability & Storage:** Keep as concentrated solution. Store at 2~8°C and protected from prolonged exposure to light. Do not freeze. Centrifuge before opening to ensure complete recovery of vial contents. This product is guaranteed up to one year from purchase.
- **Recommended** Usage: Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 μ L of antibody per test (million cells in 100 μ L staining volume or per 100 μ L of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use.