

Product Datasheet **GenieFluor 488 Anti-Human CD24 Antibody [ML5]** Catalogue Code: AGEL2271

## Antibody Data

Product SKU:	AGEL2271	Clone:	ML5
Applications:	FCM		
Reactivity:	Human		

## **Important Note:**

Centrifuge before opening to ensure complete recovery of vial contents.

## **Product Information:**

Alternate Names:	GPI linked surface mucin;Heat stable antigen;HSA;Nectadrin;Signal transducer CD24;			
Uniprot ID:	P25063			
Background:	CD24 is a 35-45 kD glycosylphosphatidylinositol (GPI)-linked protein also known as heat stable antigen (HSA), BA-1, Ly-52, and nectadrin. It is expressed on the surface of B cells (but not plasma cells), granulocytes, follicular dendritic cells, and epithelial cells. CD24 may play a role in the regulation of B-cell proliferation and maturation. CD24 crosslinking induces a Ca2+ flux in mature B cells. CD24 has been shown to interact with CD62P (P-selectin).			
Form:	Liquid	488 Excitation and Emission Spectra		
Conjugation:	Genie Fluor488	100		
Size:	20 Tests, 100 Tests, 200 Tests	80 - S		
Host Species:	Mouse	(%) 00 00 00 00 00 00 00 00 00 00 00 00 00		
Isotype:	Mouse IgG2a, к	E 40 20 0 350 400 450 500 550 600 650 700 Wavelength (nm)		

**Isotype Control:** Genie Fluor 488 Mouse IgG2a, κ Isotype Control[C1.18.4] [Product AGEL2271]

**Storage Buffer:** Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer and 1% protein protectant.

Shipping: Biological ice pack at 4°C



- **Stability & Storage:** Keep as concentrated solution. Store at 2~8°C and protected from prolonged exposure to light. Do not freeze. Centrifuge before opening to ensure complete recovery of vial contents. This product is guaranteed up to one year from purchase.
- **Recommended** Usage: Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5  $\mu$ L of antibody per test (million cells in 100  $\mu$ L staining volume or per 100  $\mu$ L of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use.