

Product Datasheet **PE/Cyanine5 Anti-Human CD28 Antibody** [**CD28.2**] Catalogue Code: AGEL2370

Antibody Data

Product SKU:	AGEL2370	Clone:	CD28.2
Applications:	FCM		
Reactivity:	Human		

Important Note:

Centrifuge before opening to ensure complete recovery of vial contents.

Product Information:

Alternate Names:	PAS2;MNS;HGpMiX;GYPA;GPSAT	PAS2;MNS;HGpMiX;GYPA;GPSAT;		
Uniprot ID:	P10747			
Background:	CD28 is a 44 kD disulfide-linked homodimeric type I glycoprotein. It is a member of the immunoglobulin superfamily and is also known as T44 or Tp44. CD28 is expressed on most T lineage cells, NK cell subsets, and plasma cells. CD28 binds both CD80 and CD86 using a highly conserved motif MYPPY in the CDR3-like loop. CD28 is considered a major co-stimulatory molecule, inducing T lymphocyte activation and IL-2 synthesis, and preventing cell death. In vitro studies indicate that ligation of CD28 on T cells by CD80 and CD86 on antigen presenting cells provides a costimulatory signal required for T cell activation and proliferation.			
Form:	Liquid	PE/Cyanine5 Excitation and Emission Spectra		
Conjugation:	PE/Cyanine 5	100		
Size:	20 Tests, 100 Tests, 200 Tests	80 - 8		
Host Species:	Mouse	(%) 60 pszirem 40		
Isotype:	Mouse IgG1, к			
		335 435 435 556 556 666 656 766 756 666 Wavelength (nm)		
		Ex:495;565;655 nm; Em:670 nm		
Isotype Control:	-			
Storage Buffer:	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer and 1% protein protectant.			

Shipping: Biological ice pack at 4°C



- **Stability & Storage:** Keep as concentrated solution. Store at 2~8°C and protected from prolonged exposure to light. Do not freeze. Centrifuge before opening to ensure complete recovery of vial contents. This product is guaranteed up to one year from purchase.
- **Recommended** Usage: Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 μ L of antibody per test (million cells in 100 μ L staining volume or per 100 μ L of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use.