



Technical Manual

AMOZ (Nitrofurantoin Furoxan) ELISA Kit

- **Catalogue Code: FSES0002**
- **Competitive ELISA Kit**
- **Research Use Only**

Contents

1. Key features and Sample Types	3
2. Storage	3
3. Test Principle.....	3
4. Kit Contents.....	4
5. Experimental Preparation	5
6. Assay Procedure.....	7
7. Data Analysis.....	7
8. Notes.....	8

1. Key features and Sample Types

Sensitivity:

0.05 ppb (ng/mL)

Assay Procedure

25°C, 45 min~15 min

Detection Limit:

Muscle, Liver, Honey, Milk, Egg - 0.1 ppb;
Milk powder, Egg powder, Feed - 0.1 ppb

Cross Reactivity:

Nitrofurantoin - 100%, Nitrofurantoin metabolite - < 0.1%, Aminohydantoin Hydrochloride - < 0.1%, Nitrofurazone metabolite - < 0.1%

Sample Recovery rate:

Muscle, Liver, Egg - 80%±25%, Honey, Milk - 75%±15%,
Milk powder, Egg powder, Feed - 85%±25%

Storage:

2-8°C for 6 months.

Expiry:

See Kit Label

2. Storage

Store the kit at 2-8°C. Do not freeze any test kit components. Return any unused microwells to their original foil bag and reseal them together with the desiccant provided and further store at 2 - 8°C.

3. Test Principle

This kit uses a Competitive-ELISA method. It can detect Nitrofurantoin (AMOZ) in samples, such as honey, muscle, etc. This kit is composed of ELISA Microtiter plate, HRP conjugate, antibody working solution, standard and other supplementary reagents. The microtiter plate in this kit has been pre-coated with coupled antigen. During the reaction, AMOZ in the samples or standard competes with coupled antigen on the solid phase supporter for sites of anti- AMOZ antibody. Then Horseradish Peroxidase (HRP) conjugate is added to each microtiter plate well, and substrate reagent is added for colour development. There is a negative correlation between the OD value of samples and the concentration of AMOZ. The concentration of AMOZ in the samples can be calculated by comparing the OD of the samples to the standard curve.

4. Kit Contents

Each kit contains reagents for 96 assays including:

No.	Component	96-WellKit
1	ELISA Microtiter plate	96 wells
2	Standards	1 mL each (0 ppb, 0.05 ppb, 0.15 ppb, 0.45 ppb, 1.35 ppb, 4.05 ppb)
3	Derivatization Reagent	10 mL
4	HRP Conjugate	5.5 mL
5	Antibody Working Solution	5.5 mL
6	Substrate Reagent A	6 mL
7	Substrate Reagent B	6 mL
8	Stop Solution	6 mL
9	20×Concentrated Wash Buffer	40 mL
10	2×Reconstitution Buffer	50 mL
11	Plate Sealer	3 pieces
12	Sealed Bag	1 piece
13	Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution.

Additional materials required:

Other materials required but not supplied

- **Instruments:** Microplate reader, Printer, Homogenizer, Nitrogen evaporators, Water bath, Vortex mixer, Centrifuge, Graduated pipette, Balance (sensitivity 0.01 g).
- **Micropipette:** Single channel (20-200 μ L, 100-1000 μ L), Multichannel (30-300 μ L).
- **Reagents:** Ethyl acetate, N-hexane, NaOH, HCl, $K_2HPO_4 \cdot 3H_2O$, $ZnSO_4 \cdot 7H_2O$, Concentrated HCl $Na_2Fe(CN)_5 NO \cdot 2H_2O$.

5. Experimental Preparation

Bring all reagents and samples to room temperature before use.

Open the micro-plate reader in advance, preheat the instrument, and set the testing parameters.

1. Sample pre-treatment Notice:

Experimental apparatus should be clean, and the pipette should be disposable to avoid cross-contamination during the experiment.

2. Solution preparation

Solution 1: 0.36 M $\text{Na}_2\text{Fe}(\text{CN})_5\text{NO} \cdot 2\text{H}_2\text{O}$ Solution (for milk, milk powder, egg powder sample)

Dissolve 10.7g of $\text{Na}_2\text{Fe}(\text{CN})_5\text{NO} \cdot 2\text{H}_2\text{O}$ to 100 mL with deionized water.

Solution 2: 1.04 M ZnSO_4 Solution (for milk, milk powder, egg powder sample)

Dissolve 29.8g of $\text{ZnSO}_4 \cdot 7\text{H}_2\text{O}$ to 100 mL with deionized water.

Solution 3: 0.1 M K_2HPO_4 Solution

Dissolve 11.4g of $\text{K}_2\text{HPO}_4 \cdot 3\text{H}_2\text{O}$ to 500 mL with deionized water.

Solution 4: 1 M HCl Solution

Dilute 8.6 mL of **Concentrated HCl** to 100 mL with deionized water.

Solution 5: 1 M NaOH Solution

Dissolve 4g of **NaOH** to 100 mL with deionized water.

Solution 6: Reconstitution Buffer

Dilute the **2xReconstitution Buffer** with deionized water.

(**2xReconstitution Buffer** (V): Deionized water (V) = 1:1). The

Reconstitution buffer can be store at 4°C for 1 month.

Solution 7: Wash Buffer

Dilute **20xConcentrated Wash Buffer** with deionized water.

(**20xConcentrated Wash Buffer** (V): Deionized water (V) = 1:19).

3. Sample pre-treatment

Targets may be distributed unevenly, resulting in no detection. To avoid this, ensure to take sufficient samples when sampling.

3.1 Pre-treatment of milk sample:

1. Take 5 mL of milk into 50 mL centrifuge tube, add 250 μL of **0.36 M $\text{Na}_2\text{Fe}(\text{CN})_5\text{NO} \cdot 2\text{H}_2\text{O}$ Solution** (Solution 1) and vortex for 30s, then add 250 μL of **1.04 M ZnSO_4 Solution** (Solution 2) and vortex for 30s, centrifuge at 4000 r/min for 10min at 15°C.
2. Take 1.1 mL of supernatant, add 4 mL of deionized water, 0.5 mL of **1 M HCl Solution** (Solution 4) and 100 μL of **Derivatization Reagent**, vortex for 5 min.
3. Incubate overnight at 37°C (about 16 hours) or incubate in water bath at 50°C for 3 hours (the effect of stratification will be affected when more than 50°C).
4. Add 5 mL of **0.1 M K_2HPO_4 Solution** (Solution 3), 0.4 mL of **1 M NaOH Solution** (Solution 5) and 5 mL of **Ethyl acetate**, vortex for 5 min.
5. Centrifuge at 4000 r/min at room temperature for 10 min.
6. Take 2.5 mL of upper liquid to another tube, dry at 50-60°C with nitrogen

evaporators or water bath.

7. Dissolve the residual with 1 mL **N-hexane**, add 1 mL of **Reconstitution Buffer** (Solution 6) and vortex for 30s. Centrifuge at 4000 r/min at room temperature for 10 min.
8. Discard the upper n-hexane, take 50 μ L lower liquid for analysis.
Note: Sample dilution factor: 2, detection limit: 0.1 ppb

3.2 Pre-treatment of milk powder, egg powder sample:

1. Weigh 1 ± 0.05 g of sample into 50 mL EP tube, add 4 mL of deionized water, 0.5 mL of **1 M HCl Solution** (Solution 4) and 100 μ L of **Derivatization Reagent**, vortex for 5 min.
2. Incubate overnight at 37°C (about 16 hours) or incubate with water bath at 50°C for 3 hours (the effect of stratification will be affected when more than 50°C).
3. Add 250 μ L of **0.36 M Na₂Fe (CN)₅ NO•2H₂O Solution** (Solution 1), vortex for 30s, then add 250 μ L of **1.04 M ZnSO₄ Solution** (Solution 2), vortex for 30s centrifuge at 4000 r/min at 15°C for 10 min.
4. Take the supernatant to another tube, add 5mL of **0.1 M K₂HPO₄ Solution** (Solution 3), 0.4 mL of **1 M NaOH Solution** (Solution 5) and 5 mL of **Ethyl acetate**, vortex for 5 min.
5. Centrifuge at 4000 r/min at room temperature for 10 min.
6. Take 2.5 mL of upper liquid to another tube, dry at 50-60°C with nitrogen evaporators or water bath.
7. Dissolve the residual with 1mL **N-hexane**, add 1 mL of **Reconstitution Buffer** (Solution 6) and vortex for 30s. Centrifuge at 4000 r/min at room temperature for 10 min.
8. Discard the upper n-hexane, take 50 μ L of lower liquid for analysis.
Note: Sample dilution factor: 2, detection limit: 0.1 ppb

3.3 Pre-treatment of honey, muscle (livestock, fish, shrimp), liver, feed, egg sample:

1. Weigh 1 ± 0.05 g of homogenate sample into 50 mL centrifuge tube, add 4 mL of deionized water, 0.5 mL of **1 M HCl Solution** (Solution 4) and 100 μ L of **Derivatization Reagent**, vortex for 5min.
2. Incubate overnight at 37°C (about 16 hours) or incubate in water bath at 50°C for 3 hours (the effect of stratification will be affected when more than 50°C).
3. Add 5 mL of **0.1 M K₂HPO₄ Solution** (Solution 3), 0.4 mL of **1 M NaOH Solution** (Solution 5) and 5 mL of **Ethyl acetate**, vortex for 5 min.
4. Centrifuge at 4000 r/min at room temperature for 10 min.
5. Take 2.5 mL of upper liquid to another tube, dry at 50-60°C with nitrogen evaporators/water bath.
6. Dissolve the residual with 1 mL **N-hexane**, add 1 mL of **Reconstitution Buffer**

(Solution 6) and vortex for 30s. Centrifuge at 4000 r/min at room temperature for 10 min.

7. Discard the upper n-hexane, take 50 µL lower liquid for analysis.

Note: Sample dilution factor: 2, detection limit: 0.1ppb

6. Assay Procedure

Bring all reagents and samples to room temperature (25°C) before use. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming. The unused ELISA Microtiter plate should be sealed as soon as possible and stored at 2~8°C.

1. **Number:** number the sample and standard in order (multiple well), and keep a record of standard wells and sample wells. **Standard and Samples must be tested in duplicate.**
2. **Add Sample:** add 50 µL of **Standard or Sample** per well, then add 50 µL of **HRP Conjugate** to each well, then add 50 µL of **Antibody Working Solution**, cover the plate with plate sealer, vortex for 5s and mix thoroughly, incubate at 25°C for 45 min in the dark.
3. **Wash:** uncover the sealer carefully, remove the liquid in each well. Immediately add 300 µL of **Wash Buffer** (Solution 7) to each well and wash. Repeat wash procedure for 5 times, 30s intervals/time. Invert the plate and pat it against absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
4. **Colour Development:** add 50 µL of **Substrate Reagent A** to each well, and then add 50 µL of **Substrate Reagent B**. Gently vortex for 5s to mix thoroughly. Incubate at 25°C for 15 min in the dark (The reaction time can be extended according to the actual colour change).
5. **Stop Reaction:** add 50 µL of **Stop Solution** to each well, vortex gently to mix thoroughly.
6. **OD Measurement:** determine the optical density (OD value) of each well at 450 nm (reference wavelength 630 nm) with a microplate reader. This step should be finished in 10 min after stop reaction.

7. Data Analysis

1. **Absorbance (%) = $A/A_0 \times 100\%$**

A: Average absorbance of standard or sample

A₀: Average absorbance of 0 ppb Standard

2. **Drawing and calculation of standard curve**

Create a standard curve by plotting the absorbance percentage of each standard on the y-axis against the log concentration on the x-axis to draw a semi-logarithmic plot. Add average absorbance value of sample to standard curve to get corresponding concentration. **If samples have been diluted, the concentration calculated from the standard curve must be multiplied by the dilution factor.**

For this kit, it is more convenient to use professional analysis form for accurate and fast analysis on a large number of samples.

8. Notes

1. The overall OD value will be lower when reagents have not been brought to room temperature before use or room temperature is below 25°C.
2. If the wells turn dry during the washing procedure, it will lead to bad linear standard curve and poor repeatability. Operate the next step immediately after wash.
3. Mix thoroughly and wash the plate completely. The consistency of wash procedure can strongly affect the reproducibility of this ELISA kit.
4. ELISA Microtiter plate should be covered by plate sealer. Avoid the kit to strong light.
5. **Each reagent is optimized for use in the FSES0002. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other FSES0002 with different lot numbers.**
6. Substrate Reagent should be abandoned if it turns blue colour. When OD value of standard (concentration: 0) < 0.5 unit (A450nm < 0.5), it indicates the reagents are deteriorated.
7. Stop solution is caustic, avoid contact with skin and eyes.
8. As the OD values of the standard curve may vary according to the conditions of the actual assay performance (e.g. operator, pipetting technique, washing technique or temperature effects), the operator should establish a standard curve for each test.
9. Even the same operator might get different results in two separate experiments. In order to get reproducible results, the operation of every step in the assay should be controlled.
10. If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
11. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.

Assay Genie 100% money-back guarantee!

If you are not satisfied with the quality of our products and our technical team cannot resolve your problem, we will give you 100% of your money back.

Contact Details



Email: info@AssayGenie.com

Web: www.AssayGenie.com