



## Technical Manual

# Human/Mouse/Rat Phospho-AKT (S473) and Total Akt PharmaGenie ELISA Kit

- **Catalogue Code: SBRS1745**
- **Sandwich Principle**
- **Research Use Only**

---

## Contents

1. Key Features and Sample Types.....	3
2. Introduction .....	4
3. Storage and Expiry.....	4
4. Kit Contents.....	5
5. Additional Materials Required .....	5
6. Reagent Preparation .....	6
7. Assay Procedure .....	7
8. Assay Procedure Summary .....	8
9. Typical Results.....	8
10. Troubleshooting Guide .....	10

---

---

## Key features and Sample Types

### Aliases:

RAC-alpha serine/threonine-protein kinase (EC 2.7.11.1) (Protein kinase B) (PKB) (Protein kinase B alpha) (PKB alpha) (Proto-oncogene c-Akt) (RAC-PK-alpha)

### Targets Detected:

2

### Gene ID:

391

### Gene Names:

AKT1 / PKB / RAC

### Uniprot:

P31749

### Detection method:

Sandwich (Semi-Quantitative)

### Sample Types:

Cell Lysates/Tissue Lysates

### Pathway:

Insulin Signaling, mTOR Signaling, PI3K-AKT Signaling

### Reactivity:

Human, Mouse, Rat

---

## Storage & Expiry

The entire kit may be stored at -20°C for up to 6 months from the date of shipment. Avoid repeated freeze-thaw cycles. For extended storage, it is recommended to store at -80°C. For prepared reagent storage, see table below.

---

## Introduction

### How do our ELISA kits work?

The Assay Genie Phospho-AKT (Ser473) and Total AKT ELISA kit is a very rapid, convenient and sensitive assay kit that can monitor the activation or function of important biological pathways in Human, Mouse and Rat cell lysates. By determining phosphorylated AKT protein in your experimental model system, you can verify pathway activation in your cell lysates. You can simultaneously measure numerous different cell lysates without spending excess time and effort in performing a Western Blotting analysis.

This Sandwich ELISA kit is an in vitro enzyme-linked immunosorbent assay for the measurement of Human, Mouse and Rat phospho-AKT and total AKT. An anti-pan AKT antibody has been coated onto a 96-well plate. Samples are pipetted into the wells and AKT present in a sample is bound to the wells by the immobilized antibody and the wells are washed. In select wells, rabbit anti-phospho-Akt (Ser473) antibody is added to detect phosphorylated AKT. In the remaining wells, rabbit anti-pan-Akt antibody is used to detect pan AKT. After washing away unbound antibody, HRP-conjugated anti-rabbit IgG is pipetted into the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of AKT (Ser473) or pan AKT bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm.

## Kit Contents

Each kit contains reagents for 96 assays including:

No.	Component	96-Well Kit	Storage
1	Microplate coated with anti-pan-AKT	8 x 12	1 month at -20°C*
2	Wash Buffer Concentrate (20X)	25ml	1 month at 4°C*
3	Positive Control-3T3S001-1	1 vial	1 week at -80°C
4	Phospho Detection Antibody anti-phospho-Akt (Ser473)	1 vial	5 days at 4°C
5	Pan Detection Antibody AKT	1 vials	5 days at 4°C
6	HRP-conjugated anti-rabbit IgG concentrate (500X)	1 vial (25µl)	Do not store and reuse.
7	TMB One-Step Substrate Reagent	12ml	N/A
8	Stop Solution	8ml	N/A
9	Assay Diluent ( 5X)	15ml	1 month at 4°C
10	Cell Lysate Buffer (2X)	5ml	1 month at 4°C

\*Return unused wells to the pouch containing desiccant pack, reseal along entire edge.

### Additional materials required:


1. Microplate reader capable of measuring absorbance at 450 nm.
2. Protease and Phosphatase inhibitors.
3. Shaker.
4. Precision pipettes to deliver 2 µl to 1 ml volumes.
5. Adjustable 1-25 ml pipettes for reagent preparation.
6. 100 ml and 1 liter graduated cylinders.
7. Absorbent paper.
8. Distilled or deionized water.
9. Log-log graph paper or computer and software for ELISA data analysis.
10. Tubes to prepare the positive control or sample dilutions.

## Sample Preparation

**Cell Lysate Preparation:** Rinse the cells with PBS, making sure to remove any remaining PBS before adding the lysis buffer. Solubilize cells at  $4 \times 10^7$  cells/ml in prepared Cell Lysate Buffer (see Reagent Preparation step 3). Pipette up and down to resuspend the pellet. Incubate the lysates with shaking at 2-8°C for 30 minutes. Microcentrifuge at 13,000 rpm for 10 minutes at 2-8°C and transfer the supernatantes into a clean test tube. Lysates should be used immediately or aliquoted and stored at -70°C. Avoid repeated freeze-thaw cycles. Thawed lysates should be kept on ice prior to use.

For the initial experiment, we recommend a serial dilution, such as a 5-fold to 50- fold dilution, for your cell lysates with prepared Assay Diluent (see Reagent Preparation step 2) before use.


**Note:** *The fold dilution of sample used depends on the abundance of phosphorylated proteins and should be determined empirically. More of the sample can be used if signals are too weak. If signals are too strong, the sample can be diluted further*



## Reagent Preparation

1. Bring all reagents and samples to room temperature (18 - 25°C) before use
2. 5X Assay Diluent should be diluted 5-fold with deionized or distilled water before use.
3. Cell lysate buffer should be diluted 2-fold with deionized or distilled water (for cell lysate and tissue lysate). We also recommend the addition of protease and phosphatase inhibitors (not included) to the lysis buffer prior to use.
4. Preparation of Positive Control: Briefly spin the Positive Control Vial. Add 500  $\mu$ l of prepared 1X Assay Diluent into Positive Control to prepare a Positive Control (P-1) solution. Gently mix the powder to allow it to dissolve thoroughly. If a precipitate is seen in the solution after mixing, this can be removed by a quick centrifuge of the positive control vial, and then pipetting the supernate only for the assay. Pipette 300  $\mu$ l 1X Assay Diluent into each tube. Use the Positive Control (P-1) solution to produce a dilution series (shown below). Mix each tube thoroughly before the next transfer. 1X Assay Diluent serves as the blank (P-0).

### DILUTION SERIES



---

5. If the Wash Concentrate (20X) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 ml of Wash Buffer Concentrate into deionized or distilled water to yield 400 ml of 1X Wash Buffer.

6. Preparation of Detection Antibodies:

**a.** Preparation of rabbit anti-phospho-Akt (Ser473) antibody: Briefly spin the vial of rabbit anti-phospho-Akt (Ser473). Add 100  $\mu$ l of 1X Assay Diluent into the vial to prepare a phospho detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days or at -80°C for one month). The concentrate should then be diluted 55-fold with 1X Assay Diluent and used in step 4 of the Assay Procedure.

**b.** Preparation of rabbit anti-pan-Akt antibody: Briefly spin the vial of rabbit anti-pan-Akt. Add 100  $\mu$ l of 1X Assay Diluent into the vial to prepare a pan detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days or at -80°C for one month). The concentrate should then be diluted 55-fold with 1X Assay Diluent and used in step 4 of the Assay Procedure.

7. Preparation of HRP-conjugated concentrates: Briefly spin the vial of HRP-conjugated anti-rabbit IgG concentrate before use. HRP-conjugated anti-rabbit IgG should be diluted 500-fold with 1X Assay Diluent and used in step 7 of the Assay Procedure.

*For example: Briefly spin the HRP-conjugated and pipette up and down to mix gently. Add 10  $\mu$ l of HRP-conjugated anti-rabbit IgG concentrate into a tube with 5 ml 1x Assay Diluent to prepare a 500-fold diluted HRP-conjugated anti-rabbit IgG solution.*



---

## Assay Procedure

1. Bring all reagents and samples to room temperature (18 - 25°C) before use. It is strongly recommended to run all positive controls and samples in at least duplicate. It is also recommended to run the positive controls in singlet for each of the pan and phospho-specific antibodies.
2. See plate layout (page 6) and label removable 8-well strips as appropriate for your experiment.
3. Add 100 µl of positive control (see Reagent Preparation step 4) or sample into appropriate wells. Cover the wells and incubate for 2.5 hours at room temperature or overnight at 4°C with gentle shaking.
4. Discard the solution and wash 4 times with 1X Wash Solution. Wash by filling each well with Wash Buffer (300 µl) using a multi-channel Pipette or auto-washer. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 µl of prepared 1X rabbit anti-phospho- AKT (Ser473) antibody (see Reagent Preparation step 6a) into the wells designated to detect phosphorylated protein. Add 100 µl of prepared 1X rabbit anti-pan-Akt antibody (see Reagent Preparation step 6b) to the remaining wells to detect pan protein. Incubate for 1 hour at room temperature with gentle shaking.
6. Discard the solution. Repeat the wash as in step 4.
7. Add 100 µl of prepared HRP-conjugated anti-rabbit IgG solution (see Reagent Preparation step 7a) to each well. Incubate for 1 hour at room temperature with gentle shaking.
8. Discard the solution. Repeat the wash as in step 4.
9. Add 100 µl of TMB One-Step Substrate Reagent to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking.
10. Add 50 µl of Stop Solution to each well. Read at 450 nm immediately.

## Assay Procedure Summary


1. Prepare all reagents, samples and positive control as instructed.
2. Add 100  $\mu$ l positive control or sample to each well. Incubate 2.5 hours at room temperature or overnight at 4°C with gentle shaking.
3. Add 100  $\mu$ l prepared detection antibody to each well. Incubate for 1 hour at room temperature with gentle shaking.
4. Add 100  $\mu$ l prepared HRP-Conjugated solution. Incubate for 1 hour at room temperature with gentle shaking.
5. Add 100  $\mu$ l TMB One-Step Substrate Reagent to each well. Incubate 30 minutes at room temperature.
6. Add 50  $\mu$ l Stop Solution to each well. Read at 450 nm immediately.

## Typical Data

Calculate the mean absorbance for each sample. Then, subtract the average zero (blank) optical density from each sample mean and set of singlet positive controls.

### A. Positive Control


NIH3T3 cells were treated with PDGFBB at 37°C for 10 min. Cells were solubilized at  $4 \times 10^7$  cells/ml in Cell Lysate Buffer. Serial dilutions of lysates were analyzed in this ELISA (see Reagent Preparation step 4).




## B. Recombinant Human PDGF Stimulation of NIH3T3 Cell

NIH3T3 cells were untreated or treated with recombinant human PDGF for 10 min. Cell lysates were analyzed using this phospho ELISA and Western Blot.

### i. ELISA




### ii. Western-Blot Analysis



### C. Sensitivity

The NIH3T3 cells were treated with recombinant human PDGF for 10 minutes to induce phosphorylation of Akt. Serial dilutions of lysates were analyzed in this ELISA and by Western blot. Immunoblots were incubated with anti-phospho-Akt (Ser473).

#### i. ELISA



## Troubleshooting

Problem	Causes	Solutions
Low signal in samples	<ul style="list-style-type: none"> <li>• Sample concentration is too low</li> <li>• Improper preparation of detection antibody</li> <li>• Too brief incubation times</li> <li>• Inadequate reagent volumes or improper dilution</li> </ul>	<ul style="list-style-type: none"> <li>• Increase sample concentration</li> <li>• Briefly spin down vials before opening. Dissolve the powder thoroughly.</li> <li>• Ensure sufficient incubation time; assay procedure step 3 may be done overnight</li> <li>• Check pipettes and ensure correct preparation</li> </ul>
High signal in samples	<ul style="list-style-type: none"> <li>• Sample concentration too high</li> </ul>	<ul style="list-style-type: none"> <li>• Reduce sample concentration</li> </ul>
Large CV	<ul style="list-style-type: none"> <li>• Inaccurate pipetting</li> <li>• Air bubbles in wells</li> </ul>	<ul style="list-style-type: none"> <li>• Check pipettes</li> <li>• Remove bubbles in wells</li> </ul>
High background	<ul style="list-style-type: none"> <li>• Plate is insufficiently washed</li> <li>• Contaminated wash buffer</li> </ul>	<ul style="list-style-type: none"> <li>• Review the manual for proper wash. If using a plate washer, ensure that all ports are unobstructed.</li> <li>• Make fresh wash buffer</li> </ul>
Low sensitivity	<ul style="list-style-type: none"> <li>• Improper storage of the ELISA kit</li> <li>• Stop solution</li> <li>• Improper primary or secondary antibody dilution</li> </ul>	<ul style="list-style-type: none"> <li>• Store your standard at <math>&lt;-70^{\circ}\text{C}</math> after reconstitution, others at <math>4^{\circ}\text{C}</math>. Keep substrate solution protected from light.</li> <li>• Add stop solution to each well before reading plate</li> <li>• Ensure correct dilution</li> </ul>

---

### **Assay Genie 100% money-back guarantee!**

If you are not satisfied with the quality of our products and our technical team cannot resolve your problem, we will give you 100% of your money back.

### **Contact Details**



Email: [info@ASSAYGenie.com](mailto:info@ASSAYGenie.com)

Web: [www.ASSAYGenie.com](http://www.ASSAYGenie.com)