

GenieFluor 488 Anti-Mouse CD3epsilon Antibody [145-2C11]

AGEL0377

Description

This GenieFluor 488 Anti-Mouse CD3epsilon Antibody [145-2C11] is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

Product Information

SKU:	AGEL0377
Contents:	100 Tests, 200 Tests, 50 Tests Bradford Reagent: 1 vial (2ml)
Category:	Monoclonal Antibody
Clonality:	Monoclonal
Clone:	145-2C11
Synonyms:	CD3E, CD3e, T-cell surface antigen T3/Leu-4 epsilon chain, T-cell surface glycoprotein CD3 epsilon chain, T3E
Applications:	FCM
Reactivity:	Mouse
Immunogen:	-

Antibody Data

Uniprot ID:	P22646
Gene ID:	12501
Swissprot:	P22646
Host Species:	Armenian Hamster
Isotype:	Armenian Hamster IgG
Isotype Control:	-

Manufacturers Statement: This final kit system is assembled and quality-released by Assay Genie Limited.

Conjugation:	GenieFluor488
Conjugation Information:	GenieFluor 488 is designed to be excited by the Blue laser (488 nm) and detected using an optical filter centered near 520 nm (e.g., a 525/40 nm bandpass filter).
Buffer:	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer.
Purification:	-
Target:	CD3εμ
Cellular Localization:	Membrane
Tissue Specificity:	-
Verified Samples:	-
Concentration:	5 μL/Test

Preparation & Storage

Storage: This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze.

Shipping: Ice bag

Recommended Dilution: -

Recommended Usage:

Application	Recommended Usage
FCM	Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 μL of antibody per test (million cells in 100 μL staining volume or per 100 μL of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use

Protein Quantification (Optional): To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

Notes: Centrifuge before opening to ensure complete recovery of vial contents.