

# GenieFluor 647 Anti-Mouse CD62L Antibody [Mel14]

AGEL0633

## Description

---

This GenieFluor 647 Anti-Mouse CD62L Antibody [Mel14] is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

## Product Information

---

<b>SKU:</b>	AGEL0633
<b>Contents:</b>	100 Tests, 200 Tests, 50 Tests Bradford Reagent: 1 vial (2ml)
<b>Category:</b>	Monoclonal Antibody
<b>Clonality:</b>	Monoclonal
<b>Clone:</b>	MEL-14
<b>Synonyms:</b>	CD62L, CD62 antigen-like family member L, L-selectin, LAM-1, LECAM1, Lnh, Ly-22, Ly22, Lymph node homing receptor, Sell
<b>Applications:</b>	<a href="#">FCM</a>
<b>Reactivity:</b>	Mouse
<b>Immunogen:</b>	-

## Antibody Data

---

<b>Uniprot ID:</b>	P18337
<b>Gene ID:</b>	20343
<b>Swissprot:</b>	P18337
<b>Host Species:</b>	Rat
<b>Isotype:</b>	Rat IgG2a, κ
<b>Isotype Control:</b>	GenieFluor 647 Rat IgG2a, κ Isotype Control[2A3]

**Manufacturers Statement:** This final kit system is assembled and quality-released by Assay Genie Limited.

<b>Conjugation:</b>	GenieFluor647
<b>Conjugation Information:</b>	GenieFluor 647 is designed to be excited by the Red laser (627-640 nm) and detected using an optical filter centered near 670 nm (e.g., a 660/20 nm bandpass filter).
<b>Buffer:</b>	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer.
<b>Purification:</b>	-
<b>Target:</b>	CD62L
<b>Cellular Localization:</b>	Membrane
<b>Tissue Specificity:</b>	-
<b>Verified Samples:</b>	-
<b>Concentration:</b>	5 µL/Test

## Preparation & Storage

---

**Storage:** This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze.

**Shipping:** Ice bag

**Recommended Dilution:** -

**Recommended Usage:**

Application	Recommended Usage
FCM	Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 µL of antibody per test (million cells in 100 µL staining volume or per 100 µL of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use

**Protein Quantification (Optional):** To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

**Notes:** Centrifuge before opening to ensure complete recovery of vial contents.