

# GenieFluor 488 Anti-Mouse CD117 Antibody [2B8]

AGEL2181

## Description

---

This GenieFluor 488 Anti-Mouse CD117 Antibody [2B8] is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

## Product Information

---

<b>SKU:</b>	AGEL2181
<b>Contents:</b>	100µg, 25µg Bradford Reagent: 1 vial (2ml)
<b>Category:</b>	Monoclonal Antibody
<b>Clonality:</b>	Monoclonal
<b>Clone:</b>	2B8
<b>Synonyms:</b>	CD117, Kit, Mast/stem cell growth factor receptor Kit, Proto-oncogene c-Kit, SCFR, Tyrosine-protein kinase Kit, c-Kit
<b>Applications:</b>	<a href="#">FCM</a>
<b>Reactivity:</b>	Mouse
<b>Immunogen:</b>	-

## Antibody Data

---

<b>Uniprot ID:</b>	P05532
<b>Gene ID:</b>	16590
<b>Swissprot:</b>	P05532
<b>Host Species:</b>	Rat
<b>Isotype:</b>	Rat IgG2b, κ
<b>Isotype Control:</b>	GenieFluor 488 Rat IgG2b, κ Isotype Control[LTF-2]

**Manufacturers Statement:** This final kit system is assembled and quality-released by Assay Genie Limited.

<b>Conjugation:</b>	GenieFluor488
<b>Conjugation Information:</b>	GenieFluor 488 is designed to be excited by the Blue laser (488 nm) and detected using an optical filter centered near 520 nm (e.g., a 525/40 nm bandpass filter).
<b>Buffer:</b>	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer.
<b>Purification:</b>	-
<b>Target:</b>	CD117
<b>Cellular Localization:</b>	Membrane
<b>Tissue Specificity:</b>	-
<b>Verified Samples:</b>	-
<b>Concentration:</b>	0.5 mg/mL

## Preparation & Storage

---

**Storage:** This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze.

**Shipping:** Ice bag

**Recommended Dilution:** -

**Recommended Usage:**

Application	Recommended Usage
FCM	Each lot of this antibody is quality control tested by flow cytometric analysis. Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use. We suggest each investigator should titrate the reagent to obtain optimal results [The recommended concentration is 0.1-1 µg/10 <sup>6</sup> cells in 100 µL volume].

**Protein Quantification (Optional):** To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

**Notes:** Centrifuge before opening to ensure complete recovery of vial contents.