

Anti-Human CD3-FITC/CD19-APC/CD16+CD56-PE Cocktail

AGEL2714

Description

This Anti-Human CD3-FITC/CD19-APC/CD16+CD56-PE Cocktail is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

Product Information

SKU:	AGEL2714
Contents:	100 Tests, 20 Tests, 200 Tests Bradford Reagent: 1 vial (2ml)
Category:	Monoclonal Antibody
Clonality:	Monoclonal
Clone:	OKT-3, CB19, 3G8, 5.1H11
Synonyms:	-
Applications:	FCM
Reactivity:	Human
Immunogen:	-

Antibody Data

Uniprot ID:	-
Gene ID:	-
Swissprot:	-
Host Species:	-
Isotype:	-
Isotype Control:	-

Conjugation:	FITC, APC, PE
Conjugation Information:	FITC is designed to be excited by the Blue laser (488 nm) and detected using an optical filter centered near 530 nm (e.g., a 525/40 nm bandpass filter). APC is designed to be excited by the Red (627-640 nm) laser and detected using an optical filter centered near 660 nm (e.g., a 660/20 nm bandpass filter). PE is designed to be excited by the Blue (488 nm), Green (532 nm) and Yellow-Green (561 nm) lasers and detected using an optical filter centered near 575 nm (e.g., a 585/42 nm bandpass filter).
Buffer:	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer.
Purification:	-
Target:	-
Cellular Localization:	-
Tissue Specificity:	-
Verified Samples:	-
Concentration:	5 µL/Test

Preparation & Storage

Storage:	This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze.
Shipping:	Ice bag
Recommended Dilution:	-

Recommended Usage:	Application	Recommended Usage
	FCM	For whole blood samples, add 5 µL Anti-Human CD3-FITC/CD19-APC/CD16+CD56-PE Cocktail to 100 µL anticoagulant-treated blood sample. Mix and incubate the sample at 4°C in the dark for 30 min. Remove red blood cells with RBC lysis solution following the manufacturer's instruction. Wash the cell with cell staining buffer and discard the supernatant after centrifugation at 300 g for 5 min. Resuspend the cells with 200 µL cell staining buffer and load the sample on flow cytometer for detection. For other samples, 1×10 ⁶ dissociated single cells are centrifuged at 300 g for 5 min with the supernatant discarded. Resuspend the cells

	with 100 μ L cell staining buffer and add 5 μ L Anti-Human CD3-FITC/CD19-APC/CD16+CD56-PE Cocktail. Mix and incubate the sample at 4°C in the dark for 30 min. Add cell staining buffer to each tube, centrifuge at 300 g for 5 min and discard the supernatant. Resuspend the cells with 200 μ L cell staining buffer and load the sample on flow cytometer for detection.
--	---

Protein Quantification (Optional):

To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

Notes:

Centrifuge before opening to ensure complete recovery of vial contents.