

# GenieFluor Red 780 Anti-Human CD66b Antibody [G10F5]

AGEL3388

## Description

---

This GenieFluor Red 780 Anti-Human CD66b Antibody [G10F5] is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

## Product Information

---

<b>SKU:</b>	AGEL3388
<b>Contents:</b>	100 Tests, 20 Tests, 200 Tests Bradford Reagent: 1 vial (2ml)
<b>Category:</b>	Monoclonal Antibody
<b>Clonality:</b>	Monoclonal
<b>Clone:</b>	G10F5
<b>Synonyms:</b>	CD66b, CD67, CD67 Antigen, CEACAM, CEACAM8, CGM, CGM6, Carcinoembryonic Antigen CGM, Carcinoembryonic Antigen CGM6, Carcinoembryonic antigen-related cell adhesion molecule 8, NCA, NCA-95, Non-Specific Cross-Reacting Antigen NCA
<b>Applications:</b>	<a href="#">FCM</a>
<b>Reactivity:</b>	Human
<b>Immunogen:</b>	-

## Antibody Data

---

<b>Uniprot ID:</b>	P31997
<b>Gene ID:</b>	1088
<b>Swissprot:</b>	P31997
<b>Host Species:</b>	Mouse
<b>Isotype:</b>	Mouse IgM, κ

**Manufacturers Statement:** This final kit system is assembled and quality-released by Assay Genie Limited.

<b>Isotype Control:</b>	GenieFluor Red 780 Mouse IgM, κ Isotype Control[MM-30]
<b>Conjugation:</b>	GenieFluorRed 780
<b>Conjugation Information:</b>	GenieFluor Red 780 is designed to be excited by the Red (627-640 nm) laser and detected using an optical filter centered near 770 nm (e.g., a 780/60 nm bandpass filter).
<b>Buffer:</b>	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer.
<b>Purification:</b>	-
<b>Target:</b>	CD66b
<b>Cellular Localization:</b>	Membrane
<b>Tissue Specificity:</b>	-
<b>Verified Samples:</b>	-
<b>Concentration:</b>	5 µL/Test

## Preparation & Storage

---

**Storage:** This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze.

**Shipping:** Ice bag

**Recommended Dilution:** -

**Recommended Usage:**

Application	Recommended Usage
FCM	Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 µL of antibody per test (million cells in 100 µL staining volume or per 100 µL of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use.

**Protein Quantification (Optional):** To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

**Notes:**

Centrifuge before opening to ensure complete recovery of vial contents.