

PE/Cyanine5.5 Anti-Human CD56 Antibody [B-A19]

AGEL3421

Description

This PE/Cyanine5.5 Anti-Human CD56 Antibody [B-A19] is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

Product Information

SKU:	AGEL3421
Contents:	100 Tests, 20 Tests, 200 Tests Bradford Reagent: 1 vial (2ml)
Category:	Monoclonal Antibody
Clonality:	Monoclonal
Clone:	B-A19
Synonyms:	N-CAM-1, NCAM, NCAM-1, NCAM1, Neural cell adhesion molecule 1
Applications:	FCM
Reactivity:	Human, Rhesus, Cynomolgus
Immunogen:	-

Antibody Data

Uniprot ID:	P51677
Gene ID:	4684
Swissprot:	P51677
Host Species:	Mouse
Isotype:	Mouse IgG1, κ
Isotype Control:	PE/Cyanine5.5 Mouse IgG1, κ Isotype Control[MOPC-21]

Conjugation:	PE/Cyanine 5.5
Conjugation Information:	PE/Cyanine5.5 is designed to be excited by the Blue (488 nm), Green (532 nm) and yellow-green (561 nm) lasers and detected using an optical filter centered near 690 nm (e.g., a 690/50 nm bandpass filter).
Buffer:	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer.
Purification:	-
Target:	CD56/NCAM
Cellular Localization:	Membrane
Tissue Specificity:	-
Verified Samples:	-
Concentration:	5 µL/Test

Preparation & Storage

Storage:	This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze.
Shipping:	Ice bag
Recommended Dilution:	-

Recommended Usage:	Application	Recommended Usage
	FCM	Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 µL of antibody per test (million cells in 100 µL staining volume or per 100 µL of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use

Protein Quantification (Optional): To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

Notes: Centrifuge before opening to ensure complete recovery of vial contents.