

GenieFluor 700 Anti-Human CD107a/LAMP-1 Antibody [H4A3]

AGEL4170

Description

This GenieFluor 700 Anti-Human CD107a/LAMP-1 Antibody [H4A3] is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

Product Information

SKU:	AGEL4170
Contents:	100 Tests, 20 Tests, 200 Tests Bradford Reagent: 1 vial (2ml)
Category:	Monoclonal Antibody
Clonality:	Monoclonal
Clone:	H4A3
Synonyms:	LAMP-1, LGP-120, Lysosome-Associated Membrane Protein 1, LAMP-1
Applications:	FCM
Reactivity:	Human
Immunogen:	-

Antibody Data

Uniprot ID:	P11279
Gene ID:	3916
Swissprot:	P11279
Host Species:	Mouse
Isotype:	Mouse IgG1, κ
Isotype Control:	GenieFluor 700 Mouse IgG1, κ Isotype Control[MOPC-21]

Manufacturers Statement: This final kit system is assembled and quality-released by Assay Genie Limited.

Conjugation:	GenieFluor700
Conjugation Information:	GenieFluor 700 is designed to be excited by the Red laser (627-640 nm) and detected using an optical filter centered near 719 nm (e.g., a 725/40 nm bandpass filter).
Buffer:	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer.
Purification:	-
Target:	CD107a/LAMP-1
Cellular Localization:	Membrane
Tissue Specificity:	-
Verified Samples:	-
Concentration:	-

Preparation & Storage

Storage: This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze.

Shipping: Ice bag

Recommended Dilution: -

Recommended Usage:

Application	Recommended Usage
FCM	Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 µL of antibody per test (million cells in 100 µL staining volume or per 100 µL of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use

Protein Quantification (Optional): To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

Notes: Centrifuge before opening to ensure complete recovery of vial contents.