

GenieFluor Violet 450 Anti-Mouse CD146 Antibody [ME-9F1]

AGEL4512

Description

This GenieFluor Violet 450 Anti-Mouse CD146 Antibody [ME-9F1] is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

Product Information

SKU:	AGEL4512
Contents:	50 Tests, 100 Tests Bradford Reagent: 1 vial (2ml)
Category:	Monoclonal Antibody
Clonality:	Monoclonal
Clone:	ME-9F1
Synonyms:	S-Endo 1 antigen, MUC18, MCAM, Mel-CAM, A32 antigen
Applications:	FCM
Reactivity:	Mouse
Immunogen:	-

Antibody Data

Uniprot ID:	P10810
Gene ID:	84004
Swissprot:	P10810
Host Species:	Rat
Isotype:	Rat IgG2a, κ
Isotype Control:	GenieFluor Violet 450 Rat IgG2a, κ Isotype Control[2A3]

Conjugation:	GenieFluorViolet 450
Conjugation Information:	GenieFluor Violet 450 is designed to be excited by the violet laser (405 nm) and detected using an optical filter centered near 450 nm (e.g., a 450/45 nm bandpass filter).
Buffer:	Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer.
Purification:	-
Target:	CD146
Cellular Localization:	Membrane
Tissue Specificity:	-
Verified Samples:	-
Concentration:	5 µL/Test

Preparation & Storage

Storage:	This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze.
Shipping:	Ice bag
Recommended Dilution:	-

Recommended Usage:	Application	Recommended Usage
	FCM	Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 µL of antibody per test (million cells in 100 µL staining volume or per 100 µL of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use.

Protein Quantification (Optional): To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

Notes: Centrifuge before opening to ensure complete recovery of vial contents.