

GenieFluor 700 Anti-Human CD18 Antibody [TS1/18.1.2.11]

AGEL4880

Description

This GenieFluor 700 Anti-Human CD18 Antibody [TS1/18.1.2.11] is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

Product Information

| | |
|----------------------|---|
| SKU: | AGEL4880 |
| Contents: | 20 Tests, 100 Tests Bradford Reagent: 1 vial (2ml) |
| Category: | Monoclonal Antibody |
| Clonality: | Monoclonal |
| Clone: | TS1/18.1.2.11 |
| Synonyms: | CD18, Cell surface adhesion glycoproteins LFA-1/CR3/p150+95 subunit beta, Complement receptor C3 subunit beta, Integrin beta-2, Itgb2 |
| Applications: | FCM |
| Reactivity: | Human |
| Immunogen: | - |

Antibody Data

| | |
|----------------------|---------------|
| Uniprot ID: | P05107 |
| Gene ID: | 3689 |
| Swissprot: | P05107 |
| Host Species: | Mouse |
| Isotype: | Mouse IgG1, κ |

| | |
|---------------------------------|---|
| Isotype Control: | GenieFluor 700 Mouse IgG1, κ Isotype Control[MOPC-21] |
| Conjugation: | GenieFluor700 |
| Conjugation Information: | GenieFluor 700 is designed to be excited by the Red laser (627-640 nm) and detected using an optical filter centered near 719 nm (e.g., a 725/40 nm bandpass filter). |
| Buffer: | Phosphate buffered solution, pH 7.2, containing 0.09% stabilizer. |
| Purification: | - |
| Target: | CD18 |
| Cellular Localization: | Membrane |
| Tissue Specificity: | - |
| Verified Samples: | - |
| Concentration: | - |

Preparation & Storage

| | |
|------------------------------|---|
| Storage: | This product can be stored at 2-8°C for 12 months. Please protected from prolonged exposure to light and do not freeze. |
| Shipping: | Ice bag |
| Recommended Dilution: | - |

Recommended Usage:

| Application | Recommended Usage |
|-------------|---|
| FCM | Each lot of this antibody is quality control tested by flow cytometric analysis. The amount of the reagent is suggested to be used 5 µL of antibody per test (million cells in 100 µL staining volume or per 100 µL of whole blood). Please check your vial before the experiment. Since applications vary, the appropriate dilutions must be determined for individual use |

Protein Quantification (Optional): To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol

Notes: Centrifuge before opening to ensure complete recovery of vial contents.

