Recombinant Keratin 6 Monoclonal Antibody

Description

This Recombinant Keratin 6 Monoclonal Antibody is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

Product Information

SKU:	AGEL6510
Contents:	Recombinant Keratin 6 Monoclonal Antibody (50 µL, 100 µL) Bradford Reagent: 1 vial (2ml)
Clone:	C111205
Applications:	WB, IHC, IF
Reactivity:	Human, Mouse, Rat


Antibody Data

Uniprot ID:	P48668
Size:	50 µL, 100 µL
Host Species:	Rabbit
Isotype:	Rabbit IgG
Conjugation:	Unconjugated
Buffer:	PBS, 50% glycerol, 0.05% Proclin 300, 0.05% protein protectant.


Preparation & Storage

Storage:	Store at -20°C Valid for 12 months. Avoid freeze / thaw cycles. Store Bradford Reagent at Room Temperature for 1 Year.
Shipping:	Gel Pack
Preparation:	WB 1:500-1:2000, IHC 1:200-1:1000, IF 1:50
Protein Quantification (Optional):	To quantify total protein levels, use the Bradford Reagent included in this kit. Visit https://www.assaygenie.com/bradford-protein-assay-protocol/ to view the full protocol
Notes:	Centrifuge before opening to ensure complete recovery of vial contents.


Validation Data




Immunofluorescent analysis of (100% Ice-cold methanol) fixed A431 cells using anti-Keratin 6 Monoclonal Antibody at dilution of 1:50.




Immunohistochemistry of paraffin-embedded Human breast using Keratin 6 Monoclonal Antibody at dilution of 1:1000.




Immunohistochemistry of paraffin-embedded Human esophagus using Keratin 6 Monoclonal Antibody at dilution of 1:1000.



Immunohistochemistry of paraffin-embedded Human lung squamous carcinoma using Keratin 6 Monoclonal Antibody at dilution of 1:1000.



Immunohistochemistry of paraffin-embedded Human prostate hyperplasia using Keratin 6 Monoclonal Antibody at dilution of 1:1000.



Western Blot with Keratin 6 Monoclonal Antibody at dilution of 1:2000. Lane 1: A431, Lane 2: Rat skin, Lane 3: Mouse skin Observed-MW:56 kDa Calculated-MW:60 kDa

Manufacturers Statement

This final kit system is assembled and quality-released by Assay Genie Limited.