Recombinant Pin1 Monoclonal Antibody

Description

This Recombinant Pin1 Monoclonal Antibody is supplied as a kit for advanced applications. The kit includes Bradford Reagent to quantify total protein concentration for accurate sample normalization (Optional).

Product Information

SKU:	RMEA1427
Contents:	Recombinant Pin1 Monoclonal Antibody (50 µL, 100 µL) Bradford Reagent: 1 vial (2ml)
Clone:	C171427
Applications:	WB, IF, FCM
Reactivity:	Human, Mouse, Rat


Antibody Data

Uniprot ID:	Q9QUR7
Size:	50 µL, 100 µL
Host Species:	Rabbit
Isotype:	Rabbit IgG
Conjugation:	Unconjugated
Buffer:	PBS, 50% glycerol, 0.05% Proclin 300, 0.05% protein protectant.


Preparation & Storage

Storage:	Store at -20°C Valid for 12 months. Avoid freeze / thaw cycles. Store Bradford Reagent at Room Temperature for 1 Year.
Shipping:	Gel Pack
Preparation:	WB 1:2000-1:10000, IF 1:50, FCM 1:50-1:100
Protein Quantification (Optional):	To quantify total protein levels, use the Bradford Reagent included in this kit. Visit https://www.assaygenie.com/bradford-protein-assay-protocol/ to view the full protocol
Notes:	Centrifuge before opening to ensure complete recovery of vial contents.


Validation Data




Flow cytometric analysis of mouse Pin1 expression on HeLa cells. Cells were stained with purified anti-Mouse Pin1, then a Alexa Fluor 488-conjugated second step antibody. The histogram were derived from events with the forward and side light-scatter characteristics of intact cells.




Immunofluorescent analysis of (4% Paraformaldehyde) fixed HeLa cells using anti-Pin1 Monoclonal Antibody at dilution of 1:50.



Immunofluorescent analysis of (4% Paraformaldehyde) fixed NIH-3T3 cells using anti-Pin1 Monoclonal Antibody at dilution of 1:50.



Western Blot with Pin1 Monoclonal Antibody at dilution of 1:10000. Lane 1: Mouse skin (Low expression), Lane 2: Mouse brain Observed-MW:18 kDa Calculated-MW:18 kDa



Western Blot with Pin1 Monoclonal Antibody at dilution of 1:10000. Lane 1: Mouse spleen, Lane 2: NIH/3T3, Lane 3: PC-12, Lane 4: MCF-7, Lane 5: HeLa Observed-MW:18 kDa Calculated-MW:18 kDa

Manufacturers Statement

This final kit system is assembled and quality-released by Assay Genie Limited.