



Technical Manual

Human Respiratory Syncytial Virus (HRSV) IgG ELISA Kit

- Catalogue Code: HDES0055
- Antibody ELISA Kit
- Research Use Only

1. Test principle

This ELISA kit uses Indirect-ELISA as the principle to detect the Human Respiratory Syncytial Virus (HRSV) IgG in human serum. The ELISA Microtiter plate provided in this kit has been pre-coated with purified HRSV antigen. Samples are added to the ELISA Microtiter plate wells and the HRSV-IgG in which will combine with the pre-coated antigen to form antigen-antibody compound. Free components are washed away. The HRP conjugated Mouse-anti-human IgG antibody is added to each well and react with the compound to form “antigen- antibody-HRP antibody” compound. The TMB substrate is added to initiate the color developing reaction. The presence of HRSV-IgG can be determined according to the OD value after colorimetric assay with the Micro-plate Reader.

2. Kit components

| Item | Specifications |
|---------------------------------|----------------|
| ELISA Microtiter plate | 96 wells |
| Positive Control | 1 mL |
| Negative Control | 1 mL |
| HRP Conjugated Working Solution | 12 mL |
| Sample Diluent | 12 mL |
| 20×Concentrated Wash Buffer | 50 mL |
| Substrate Reagent A | 6 mL |
| Substrate Reagent B | 6 mL |
| Stop Solution | 6 mL |
| Plate Sealer | 3 pieces |
| Sealed Bag | 1 piece |
| Manual | 1 copy |

3. Other materials required but not supplied

- Micro-plate Reader with 450 nm wavelength filter or dual-wavelength (450/630 nm)
- High-precision transferpettor, EP tubes and disposable pipette tips
- 37° C Incubator or water bath
- Deionized water
- Absorbent paper
- Loading slot for Wash Buffer

4. Notes

1. Wear gloves and work clothes during experiment, and the disinfection and isolation system should be strictly executed. All the waste should be handled as contaminant.
2. The Stop Solution is corrosive, it should be avoided to contact with skin and clothing. Wash immediately with plenty of water if contact it carelessly.
3. The ELISA Microtiter plate obtained from cold storage conditions should be adjusted to room temperature before use. The unused plate should be kept in a sealed bag with desiccant.
4. Wash Buffer at low temperature condition is easy to crystallize, it should be adjusted to room temperature to dissolve completely before use.
5. Each well must be filled with liquid when washing to prevent residual free enzyme.
6. The tested sample should be kept fresh.
7. The results shall depend on the readings of the Micro-plate Reader.
8. Do not use components from different batches of kit.

5. Storage and expiry date

Store unopened at 2-8° C. Do not freeze.

Please store the opened kit at 2-8° C, protect from light and moisture. The shelf life of the opened kit is up to 1 months.

Expiry date: expiration date is on the box.

6. Sample preparation

1. **Serum:** Human serum can be used as detected sample. Fresh collected serum samples should be fully centrifuged, then take clear liquid for test. Suspended fibrous protein may cause a false positive.
2. Avoid of samples with suspended fibrous protein, aggregation or severe hemolysis (hemoglobin > 10 g/L), hyperlipemia (triglyceride > 20 g/L), high bilirubin (bilirubin > 0.2 g/L). Obviously contaminated samples can't be detected.
3. Samples can be stored at 2-8° C for one week. If samples not tested in a week, store them below -15° C no more than a month, and avoid freeze-thaw cycles.
4. Bring all reagents to room temperature (18-25° C) for more than 30 min before use. Freezing samples should be mixed fully before test.
5. **Wash Buffer:** The **20×Concentrated Wash Buffer** should be adjusted to room temperature to make the sediment dissolved fully before use, and then dilute it with deionized water at 1:19.

7. Assay procedure

Restore all reagents and samples to room temperature (25° C) before use. All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming. The unused ELISA Microtiter plate should be sealed as soon as possible and stored at 2-8° C.

1. **Number:** number the sample and control in order (multiple well), and keep a record of control wells and sample wells. Set 1 well for blank control, 3 wells for negative control and 2 wells for positive control. **Samples need test in duplicate** (Blank well is not necessary for dual-wavelength detection).
2. **Add sample:**
 - a) Add 100 µL of **Positive/Negative Control** respectively to **Positive/Negative Control** wells, keep the blank control well empty.
 - b) Dilute the tested **Serum** with **Sample Diluent** at 1:10 into sample well (add 100 µL of sample diluent and add 10 µL of serum sample), mix fully.
 - c) Gently tap the plate to ensure thorough mixing.
3. **Incubate:** Cover the ELISA plate with sealer. Incubate for 30 min at 37° C in shading light.
4. **Wash:** remove the plate sealer and aspirate the liquid of each well. Repeat the washing procedure for 5 times with **Wash Buffer** and immerse for 30-60 sec each time. Invert the plate and pat it against thick clean absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
5. **HRP conjugate:** Add 100 µL of **HRP Conjugate Working Solution** to each well except the blank control well.
6. **Incubate:** Cover the ELISA plate with sealer. Incubate for 30 min at 37° C.
7. **Wash:** repeat step 4.
8. **Add substrate:** Add 50 µL of **Substrate Reagent A** and 50 µL of **Substrate Reagent B** to each well. Gently tap the plate to ensure thorough mixing. Cover with a new plate sealer. Incubate for 15 min at 37° C in shading light.
9. **Stop reaction:** Add 50 µL of **Stop Solution** to each well, gently tap the plate to ensure thorough mixing.
10. **OD Measurement:** Set the Micro-plate Reader wavelength at 450 nm (it is recommended to set the dual wavelength at 450 nm/630 nm) to detect A value of each well. Blank well is not needed when using dual wavelength 450 nm/630 nm for detection.

8. Reference value

Normally, Blank well (just chromogenic agent and Stop Solution) absorbance ≤ 0.08 . Positive control (PC) $A_{450} > 0.80$. Negative control (NC) $A_{450} < 0.10$.

9. Interpretation of test results

Use each test result independently. Determine the result according to the Cut Off value.

Calculate the Cut Off: Cut Off (C.O) = 0.10 + negative control (NC) average A value (when NC average $A_{450} < 0.05$, calculate at 0.05; while NC average $A_{450} \geq 0.05$, calculate at the actual value).

1. Positive result: Sample absorbance \geq Cut Off.
2. Negative result: Sample absorbance $<$ Cut Off.
3. Negative result indicates no HRSV-IgG antibody detected in samples, while positive result means the opposite. HRSV-IgG is the important indicator of HRSV infection.

10. Limitations of test method

1. This test is only used as the qualitative detection of HRSV-IgG antibody in serum of human.
2. The detection results of this kit are only for reference. For confirmation of the result, please combine the symptoms and other methods of detection, this detection cannot be used as the only criteria for result.

Assay Genie 100% money-back guarantee!

If you are not satisfied with the quality of our products and our technical team cannot resolve your problem, we will give you 100% of your money back.

Contact Details



Email: info@assaygenie.com

Web: www.assayenie.com