



## **TECHNICAL MANUAL**

# **Human NRP1 (Neuropilin 1) Quickstep ELISA Kit**

- **SKU CODE:** QSES131
- **SIZE:** 96T/48T
- **DETECTION PRINCIPLE:** Sandwich
- **RUO:** Research-Use-Only

# Human NRP1 (Neuropilin 1) Quickstep ELISA Kit

*Please read entire manual carefully before starting experiment. DO NOT mix reagents and use reagents from different kits or batches to prevent assay failure.*

## Table of Contents

<b>1. Key Features</b> .....	3
<b>2. Storage &amp; Expiry</b> .....	3
<b>3. Product Description</b> .....	4
<b>4. Kit Contents</b> .....	5
<b>5. Precautions</b> .....	6
<b>6. Assay Summary</b> .....	7
<b>7. Sample Preparation</b> .....	7
<b>7.1. Protein Quantification (Optional)</b> .....	11
<b>8. Standard and Reagent Preparation</b> .....	11
<b>9. Assay Procedure</b> .....	14
<b>10. Data Analysis</b> .....	15
<b>11. Typical Data</b> .....	16
<b>12. ELISA Troubleshooting</b> .....	18

## 1. Key Features

**Detection Method:**

Sandwich

**Sample Type:**

Serum, Plasma And Other Biological Fluids

**Reactivity:**

Human

**Range:**

0.78-50 ng/mL

**Sensitivity:**

0.47 ng/mL

## 2. Storage & Expiry

Assay Genie ELISA Kits are shipped on ice packs. Please store this ELISA Kit and/or components as described in section 4. Date of expiration is on the ELISA Box label.

### 3. Product Description

The Assay Genie Human NRP1 (Neuropilin 1) Quickstep ELISA Kit enables higher sensitivity and precision while reducing assay time by approximately 1-2 hours compared with traditional ELISA formats. This kit is designed for the rapid and accurate quantification of Human NRP1 in the following samples: Serum, Plasma And Other Biological Fluids.

This kit utilizes a sandwich enzyme-linked immunosorbent assay (ELISA) format. Each plate is pre-coated with a high-affinity antibody specific to target protein. When samples or standards are added to the wells, the target protein present binds to the immobilized antibody. A biotinylated detection antibody, also specific to target protein, is then introduced, forming a sandwich complex. This is followed by the addition of Avidin-Horseradish Peroxidase (HRP) conjugate, which binds to the biotinylated antibody. After incubation, unbound components are washed away, and a substrate solution is added. Wells containing the complete complex develop a blue color, which turns yellow upon the addition of stop solution.

The optical density is measured at  $450 \pm 2$  nm, and the intensity of the yellow color is directly proportional to the concentration of target in the sample. Concentrations are determined by comparing sample OD values to a standard curve.

**This dual function kit includes validated Bradford Reagent to quantify total protein concentration for accurate sample normalization.**

## 4. Kit Contents

An unopened kit can be stored at 2-8°C for 6 months. After opening, store the items separately according to the following conditions.

No	Component Name	Specifications	Storage
1	Micro ELISA Plate (Dismountable)	96T: 8wells× 12strips   48T: 8wells× 6strips   24T: 8wells× 3strips   96T*5: 5plates,96T	2–8°C, 1 month
2	Reference Standard	96T: 2 vials   48T/24T: 1vial   96T*5: 10 vials	2–8°C, use the reconstituted standard within 24 hours
3	Concentrated HRP Conjugate (100×)	96T: 1vial,60µL   48T/24T: 1vial,30µL   96T*5: 5vials,60µL	2–8°C (Protect from light)
4	Reference Standard & Sample Diluent	96T/48T/24T: 1vial,20mL   96T*5: 5vials,20mL	2–8°C
5	HRP Conjugate Diluent	96T/48T/24T: 1vial,14mL   96T*5: 5vials,14mL	2–8°C
6	Concentrated Wash Buffer (25×)	96T/48T/24T: 1vial,30mL   96T*5: 5vials,30mL	2–8°C
7	Substrate Reagent	96T/48T/24T: 1vial,10mL   96T*5: 5vials,10mL	2–8°C (Protect from light)
8	Stop Solution	96T/48T/24T: 1vial,10mL   96T*5: 5vials,10mL	2–8°C
9	Plate Sealer	96T/48T/24T: 5pieces   96T*5: 25pieces	2–8°C
10	Technical Manual	1 copy	-
11	Certificate of Analysis	1 copy	-
12	Bradford Reagent	96T: 1 vial   48T: 1 vial	4°C

**Additional materials required:**

1. 37°C incubator.
2. Plate Reader with 450nm filter.
3. Precision pipettes and disposable pipette tips.
4. Distilled water.
5. Disposable tubes for sample dilution.
6. Absorbent paper.

**5. Precautions**

1. This kit is for research purposes only and not for diagnostics or therapeutic uses.
2. Store all components as listed in this manual. Do not use the ELISA Kit after its expiration date.
3. Allow all reagents and samples to reach room temperature before use.
4. Ensure unopened and unused plate are kept dry to avoid contamination.
5. Before using the kit, centrifuge tubes to spin down standard and/or antibody.
6. Prepare all reagents, samples and standards as directed in this manual.
7. Duplicate wells are recommended for both standard and sample testing.
8. Do not let the microplate wells dry during the assay.
9. Maintain consistent incubation times and temperatures as variations can affect results.
10. Do not reuse tips and tubes to avoid cross contamination.
11. Avoid using the reagents from different batches together.

## 6. Assay Summary



## 7. Sample Preparation

The procedures outlined in this document are provided as general recommendations for sample preparation in ELISA assays. Due to the variability of biological samples and specific assay requirements, users are advised to optimize protocols based on their own experimental conditions.

**Note:** For information regarding validation data in specific samples, please contact our Technical Support Team at [techsupport@assaygenie.com](mailto:techsupport@assaygenie.com).

### General Considerations

To prevent denaturation or degradation of target proteins, it is recommended to process samples promptly and store them under appropriate conditions.

- **Storage Conditions:**
  - **Short-term:** 2-8 °C for up to 5 days.
  - **Medium-term:** -20 °C for up to 6 months.
  - **Long-term:** -80 °C or cryopreservation in liquid nitrogen.
- **Thawing Protocol:** Frozen samples should be thawed rapidly in a 15-25 °C water bath to minimize ice crystal-induced damage. Thawed samples can be analyzed immediately or stored temporarily at 2-8 °C.
- **Freeze-Thaw Cycles:** Repeated freeze-thaw cycles should be strictly avoided due to their detrimental effect on protein stability.

#### A. Blood-Derived Samples

- **Serum:** Allow whole blood to coagulate at room temperature (2 h) or 2-8 °C overnight. Centrifuge at 1000 × g for 20 min and collect the supernatant. Store or use immediately.
- **Plasma:** Collect in anticoagulant tubes (EDTA, citrate, or heparin), mix gently, and centrifuge within 30 min at 1000 × g, 2-8 °C for 15 min. Store or assay as needed.

#### B. Tissue Homogenates

Tissue samples should be homogenized prior to use. Avoid buffers containing NP-40, Triton X-100, or DTT, as these strongly inhibit the assay. We recommend using 50 mM Tris + 0.9% NaCl + 0.1% SDS, pH 7.3.

The recommended protocol is as follows:

- **Sample Collection and Washing**
  - Place the target tissue on ice.
  - Rinse the tissue with pre-cooled PBS buffer (0.01 M, pH 7.4) to remove residual blood.
  - Weigh the tissue for further processing.

- **Homogenization**
  - Homogenize the tissue on ice using an appropriate lysis buffer.
  - The lysate volume should correspond to the tissue weight; typically, 9 mL PBS is used per 1 g of tissue. It is recommended to add protease inhibitors to the PBS (e.g., 1 mM PMSF). **Note:** *PBS buffer or mild RIPA lysis buffer can be used for homogenization. When using RIPA, adjust pH to 7.3.*
- **Cell Disruption**
  - Further disrupt the tissue using ultrasonic homogenization or freeze–thaw cycles.
    - Ultrasonic homogenization: Keep samples on an ice bath during sonication to avoid overheating.
    - Freeze–thaw cycles: Repeat twice for effective lysis.
- **Centrifugation and Storage**
  - Centrifuge the homogenate at 5000 × g for 5 minutes.
  - Collect the supernatant for immediate analysis, or aliquot and store at –20°C or –80°C for future assays.
- **Protein Concentration Measurement**
  - Determine total protein concentration using the Bradford Reagent included in this kit.
  - For ELISA assays, the total protein concentration should generally be 1–3 mg/mL.
  - Tissues with high endogenous peroxidase levels (e.g., liver, kidney, pancreas) may react with TMB substrate, causing false positives. If this occurs, treat samples with 1% H<sub>2</sub>O<sub>2</sub> for 15 minutes before repeating the assay.

**Note:** *Liver, kidney, and pancreas samples often contain high levels of endogenous peroxidase, which may react with the chromogenic substrate at elevated sample concentrations, potentially resulting in false positive signals.*

*If analysis of these tissues is required, a gradient dilution assay is recommended. A proportional decrease in signal with increasing dilution typically indicates minimal interference and supports the accuracy of the results.*

*To further minimise potential interference, samples can be pre-treated with 1% hydrogen peroxide ( $H_2O_2$ ) for 15 minutes prior to testing. To prepare the treatment solution, add 1  $\mu$ l of pure  $H_2O_2$  to 100  $\mu$ l of sample (1% v/v).*

### **C. Cell Culture Supernatant**

Centrifuge the sample at 2500 rpm for 5 minutes at 2–8°C. Carefully collect the clarified cell culture supernatant for immediate analysis, or aliquot and store it at –80°C for future assays.

### **D. Cell Lysates**

- **Suspension Cell Lysate:** Centrifuge the cell suspension at 2500 rpm for 5 minutes at 2–8°C and collect the cell pellet. Wash the pellet with pre-cooled PBS (0.01 M, pH 7.4) and mix gently. Repeat centrifugation and discard the supernatant. Add 0.5–1 mL of cell lysis buffer containing an appropriate protease inhibitor (e.g., PMSF, final concentration: 1 mM). Lyse the cells on ice for 30–60 minutes or disrupt them using ultrasonic homogenization.
- **Adherent Cell Lysate:** Remove the supernatant and wash the cells three times with pre-cooled PBS. Add 0.5–1 mL of cell lysis buffer supplemented with an appropriate protease inhibitor (e.g., PMSF at a final concentration of 1 mmol/L). Scrape the adherent cells using a cell scraper and transfer the cell suspension to a centrifuge tube. Lyse the cells on ice for 30–60 minutes or disrupt the cells by ultrasonic treatment.

Follow next steps for protein extraction and supernatant collection:

- **Protein Release and DNA Disruption**
  - During lysis, pipette gently or intermittently shake the tube to enhance protein extraction.

- Mucilaginous material formed during lysis is DNA, which can be broken down by ultrasonic disruption (3–5 mm probe, 150–300 W, 3–5 seconds per cycle, with 30-second intervals for 1–2 minutes total).
- **Supernatant collection**
  - After lysis or ultrasonic treatment, centrifuge the lysate at 10,000 rpm for 10 minutes at 2–8°C. Collect the supernatant for immediate use or aliquot and store at –80°C for future assays.

**Notes:** Refer to the "Tissue Sample Notes" for additional buffer and inhibitor recommendations.

### **E. Other Sample Types**

For more information about how to process other sample types, (e.g., body fluids, breast milk & more), please contact our Tech Support Team at [techsupport@assaygenie.com](mailto:techsupport@assaygenie.com).

## **7.1. Protein Quantification (Optional)**

To quantify total protein levels, use the Bradford Reagent included in this kit. Visit <https://www.assaygenie.com/bradford-protein-assay-protocol/> to view the full protocol.

## **8. Standard and Reagent Preparation**

### **Manual Washing**

Discard the solution in the plate without touching the side of the wells. Clap the plate on absorbent filter paper or other absorbent material. Fill each well completely with 350 µl wash buffer and soak for 1 to 2 mins, then aspirate contents from the plate, and clap the plate on absorbent filter paper or other absorbent material. Repeat this procedure for the designated number of washes.

## Automated Washing

Aspirate all wells, then wash plate with 350  $\mu$ l wash buffer. After the final wash, invert plate, and clap the plate on absorbent filter paper or other absorbent material. It is recommended that the washer is set for a soaking time of 1 minute.

**Note:** *Set the height of the needles; be sure the fluid can be taken up completely.*

## Sample Dilution Guidelines

Determine the concentration of the target protein in the test sample and then select the optimal dilution factor to ensure the target protein concentration falls within the optimal detection range of the kit. Dilute the samples with the dilution buffer provided with the kit. Several dilution tests may be required to achieve the optimal results. The test samples must be well mixed with the dilution buffer. Standard and sample dilution should be performed before starting the experiment.

**Note:** *Dilution may be necessary to minimize matrix effects. However, if the target concentration in the sample is very low, the pre-treated sample can be added directly to the assay without dilution.*

## Reagent Preparation


Bring all reagents and samples to room temperature 20 minutes before use (18 - 25°C). For repeated assays, please use only strips and standards required and store remaining reagents at the appropriate temperatures.

### A. Wash Buffer:

Dilute 30 ml of Concentrated Wash Buffer into 750 ml of Wash Buffer with deionized or distilled water (recommended resistivity of ultrapure water is 18M $\Omega$ ). **Note:** *If crystals have formed in the concentrate, warm at 40°C in water bath (Heating temperature should not exceed 50°C) and mix gently until crystals have completely dissolved. The solution should be cooled to room temperature before use.*

**B. Standard Dilution:**

1. Centrifuge the standard tube for 1 min at 10,000 x g.
2. Add 1 mL of Reference Standard & Sample Diluent, let it stand for 10 min and invert it gently several times. After it dissolves fully, mix it thoroughly with a pipette. This reconstitution produces a working solution of 50 ng/mL. Then make serial dilutions as needed. The recommended dilution gradient is as follows: 50, 25, 12.5, 6.25, 3.13, 1.56, 0.78 and 0 ng/mL. **Note:** The final tube serves as the blank and should not receive any solution transferred from the preceding tube.
3. Take 7 Eppendorf tubes add 500  $\mu$ L of Reference Standard & Sample Diluent to each tube. Pipette 500  $\mu$ L of the 50 ng/mL working solution to the first tube and mix up to produce a 25 ng/mL working solution. Pipette 500  $\mu$ L of the solution from the former tube into the latter one according to these steps. The illustration below is for reference.



**Note:** The working solution of the standard substance after reconstitution should be aliquoted and stored at  $-20^{\circ}\text{C}$ . It should be used up within half a month and repeated freeze-thaw should be avoided. Gradient diluted standard working solution should be prepared just before use.

### C. Preparation of HRP Conjugate Working Solution:

The working solution should be prepared before starting the experiment.

1. Calculate the required amount before the experiment (50  $\mu$ L/well). **Note:** *It is advisable to prepare an amount marginally exceeding the calculated requirement.*
2. Centrifuge the Concentrated HRP Conjugate solution at 800 x g for 1 min.
3. Dilute the 100x Concentrated HRP Conjugate to 1x working solution with HRP Conjugate Diluent.

## 9. Assay Procedure

1. **Plate Setup:** Set standard, test sample and control (zero) wells on the pre-coated plate and record their positions. It is recommended to measure each standard and sample in duplicate.
2. **Standard, Samples & Control Loading:** Aliquot 50  $\mu$ L of standard working solution or samples and controls into the designated wells. Immediately add 50  $\mu$ L **HRP Conjugate Working Solution.** **Note:** *Solutions should be added to the bottom of the micro-ELISA plate well, avoid touching the inside wall and causing foaming as much as possible.*
3. **First Incubation:** Seal the plate with a cover and incubate at 37 °C for 60 mins.
4. **Washing:** Aspirate or decant the solution from the plate and add 350 $\mu$ L of wash buffer to each well and incubate for 1-2 minutes at room temperature. Aspirate the solution from each well and clap the plate on absorbent filter paper to dry. Repeat this process 3 times. **Note:** *A microplate washer can be used in this step and other wash step.*
5. **Washing:** Aspirate or decant the solution from each well. Repeat the wash process five times as conducted in step 5.
6. **Substrate Reagent Addition and Colour Development:** Add 90  $\mu$ L of Substrate Reagent into each well, cover the plate and incubate at 37°C in dark for 10-20 mins. (**Note:** *This incubation time is for reference only, the optimal time should be determined by the end-user. DO NOT exceed 30 minutes).*

7. **Stop Reaction:** Add 50  $\mu$ l of Stop solution into each well and mix thoroughly. The colour changes into yellow immediately.
8. **OD Measurement:** Read the O.D. absorbance at 450 nm in a microplate reader immediately after adding the stop solution.

## 10. Data Analysis

Average the duplicate readings for each standard, control, and sample, then subtract the mean optical density of the zero standard. Construct a standard curve by plotting Human NRP1 concentration on the y-axis against absorbance on the x-axis and fit the data using an appropriate best-fit curve.


For diluted samples, multiply the concentration obtained from the standard curve by the corresponding dilution factor to determine the final concentration. Data analysis and curve fitting may be performed using suitable plotting software (e.g., CurveExpert).

**Note:** *If the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution.*

## 11. Typical Data

### Standard Curve

Results of a typical standard run of an ELISA kit are shown below. This standard curve was generated at our lab for demonstration purpose only. Each user should obtain their own standard curve as per experiment.



### Precision

Intra-assay Precision (Precision within an assay): 3 samples with low, mid-range and high levels were tested 20 times on one plate, (n= replicate).

Inter-assay Precision (Precision between assays): 3 samples with low, mid-range and high level were tested on 3 different plates, 20 replicates in each plate.

Sample	Intra-assay Precision			Inter-assay Precision		
	1	2	3	1	2	3
n	20	20	20	20	20	20
Mean ng/mL	2.57	7.09	16.67	2.47	7.31	15.74
Standard deviation	0.11	0.36	1.05	0.2	0.49	1.29
CV (%)	4.21	5.06	6.29	8.24	6.65	8.2

## Recovery

Matrices listed below were spiked with a certain level of Human NRP1 and the recovery rates were calculated by comparing the measured value to the expected amount of Human NRP1 in the samples.

Sample Type	Range (%)	Average Recovery (%)
Serum (n=8)	94-109	100
EDTA Plasma (n=8)	92-107	99

## Linearity

The linearity of the kit was assayed by testing the samples spiked with appropriate concentration of Human NRP1 and their serial dilutions.

		Serum (n=5)	EDTA Plasma (n=5)
1:2	Range (%)	93-109	94-109
	Average (%)	101	100
1:4	Range (%)	93-104	91-102
	Average (%)	98	96
1:8	Range (%)	97-109	90-102
	Average (%)	102	97
1:16	Range (%)	97-112	95-110
	Average (%)	103	101

## 12. ELISA Troubleshooting

Problem	Possible Causes	Solutions
<b>Standard curve without signal</b>	Incorrect reagent order; Mixed components from different kits; Missing reagents.	Ensure correct reagent order and use components from the same kit. Verify all reagents are added.
<b>Overflow OD</b>	Mixed components from different kits; Over-concentrated working solution.	Use correct components and prepare solutions at recommended concentrations.
<b>Poor standard curve</b>	Incorrect curve fitting model.	Try alternative curve fitting models.
<b>Samples without signal</b>	Sample concentration too low; Incompatible buffer; Incorrect preparation; Sample degradation or excessive freeze-thaw.	Reduce dilution or concentrate sample. Check buffer compatibility and follow proper preparation and storage.
<b>High CV%</b>	Precipitate formation; Unclean plate; Foaming; Uneven washing; Incomplete reagent mixing; Pipetting inconsistency.	Dilute samples if needed, avoid foaming, ensure uniform washing, mix reagents thoroughly, and use calibrated pipettes.
<b>Low standard signal</b>	Improperly reconstituted standards; Degraded standards; Incorrect pipetting; Expired kit; Improper storage; Over-dried wells.	Reconstitute standards properly, use fresh kits, follow storage recommendations, and prevent wells from drying.
<b>Slow colour development</b>	TMB not equilibrated; Incorrect microplate reader wavelength; Over-washing.	Pre-warm TMB (30 min at 37°C), confirm correct wavelength (450 nm), and follow recommended washing times.
<b>High background</b>	Insufficient washing; Contaminated wash buffer; Excess detection reagents; Delayed reading; TMB exposed to light.	Wash adequately, prepare fresh wash buffer, use correct reagent amounts, read results promptly, and incubate TMB in the dark.

**Notes:**

**Assay Genie 100% money-back guarantee!**

If you are not satisfied with the quality of our products and our technical team cannot resolve your problem, we will give you 100% of your money back.



**Manufacturers Statement: This final kit system is assembled and quality-released by Assay Genie Limited.**